



# ISWAVLD 2023

International Symposium of the World  
Association of Veterinary Laboratory  
Diagnosticians

29 JUNE-1 JULY  
2023  
Congress Centre  
Lyon

Towards  
the veterinary  
diagnostics  
of the  
future

Main topic : One Health

## Development of one-step multiplex qPCR/RT-qPCR assays for simultaneous detection of SARS-CoV-2 and other viral and bacterial agents associated with canine and feline respiratory disease

THIEULENT C.<sup>1,2</sup>, CAROSSINO M.<sup>1,2</sup>, PEAK L.<sup>2</sup>, STROTHER K.<sup>2</sup>, LI G.<sup>3</sup>, WOLFSON W.<sup>4</sup>, BALASURIYA U.<sup>1,2</sup>

<sup>1</sup> Department of Pathobiological Sciences, School of Veterinary Medicine, Louisiana State University, Baton Rouge, United States; <sup>2</sup> Louisiana Animal Disease Diagnostic Laboratory, School of Veterinary Medicine, Louisiana State University, Baton Rouge, United States; <sup>3</sup> Department of Veterinary Diagnostic and Production Animal Medicine, College of Veterinary Medicine, Iowa State University, Ames, United States; <sup>4</sup> Department of Veterinary Clinical Sciences, School of Veterinary Medicine, Louisiana State University, Baton Rouge, United States

### Introduction

Canine infectious respiratory disease complex (CIRDC) and feline upper respiratory tract disease (URTD) are the primary causes of respiratory disease in companion animals and are associated with a wide array of viruses and bacteria acting as either individual etiologic agents or in combination, making etiologic diagnosis challenging. Additionally, SARS-CoV-2 has been reported to infect both dogs and cats. Therefore, the rapid detection and differentiation of SARS-CoV-2 from other common viral and bacterial agents in a single specimen is critical.

### Methods

Here, two panels of one-step TaqMan® multiplex qPCR/RT-qPCR were developed and validated for the identification of CIRDC and feline URTD-associated agents along with SARS-CoV-2. The canine respiratory panel included the detection of 8 viruses and 4 bacteria. The feline respiratory panel included the detection of 4 viruses and 3 bacteria. The analytical performance of each assay was evaluated using reference strains of each pathogen and plasmid DNA or *in vitro* transcribed RNA containing the target sequences. The clinical performance was evaluated using 76 and 63 nasal/nasopharyngeal swabs obtained from CIRDC-suspected dogs and URTD-suspected cats, respectively.

### Results

All the multiplex assays demonstrated high specificity, analytical sensitivity, efficiency, and linearity. Among the clinical samples derived from dogs, *Mycoplasma canis*, *M. cynos* and canine respiratory coronavirus were the most prevalent pathogens. The emerging canine pneumovirus was detected in four samples. Among the clinical samples from cats, *M. felis* was the most prevalent pathogen, followed by feline herpesvirus type-1, *Chlamydia felis* and feline calicivirus. SARS-CoV-2 was detected in four canine and two feline samples. A high prevalence of co-infection was detected, with a rate of 29% and 59% in the canine and feline respiratory samples, respectively.

### Conclusions

These two newly panels of one-step TaqMan® multiplex qPCR/RT-qPCR are useful and reliable for the rapid detection and identification of canine and feline respiratory pathogens, along with SARS-CoV-2. The high prevalence of co-infection highlights the need for the simultaneous detection of multiple pathogens using such panels in order to implement proper treatment and/or prevention plans.