



# ISWAVLD 2023

International Symposium of the World  
Association of Veterinary Laboratory  
Diagnosticians

29 JUNE-1 JULY  
2023  
Congress Centre  
Lyon

Towards  
the veterinary  
diagnostics  
of the  
future

Main topic : Other Topics

## Effect of freeze-thaw on PRRSV RNA detection by RT-qPCR

MUNGUIA-RAMIREZ B. <sup>1</sup>, ARMENTA-LEYVA B. <sup>1</sup>, HENAO-DIAZ A. <sup>2</sup>, YE F. <sup>1</sup>, DOOLITTLE K. <sup>3</sup>, ZIMMERMAN S. <sup>3</sup>, GIMENEZ-LIROLA L. <sup>1</sup>, EGLI C. <sup>3</sup>,  
TROCH J. <sup>3</sup>, ZIMMERMAN J. <sup>1</sup>

<sup>1</sup> Iowa State University, Ames, Iowa, United States; <sup>2</sup> PIC Latinoamerica, Queretaro, Queretaro, Mexico; <sup>3</sup> IDEXX Laboratories, Inc, Westbrook, Maine, United States

Re-testing samples is common in diagnostic and research settings. This implies subjecting samples to multiple freeze-thaw cycles, but what is the effect of this process on the detection of PRRSV RNA? Herein, the effect of freeze-thaw on the detection of PRRSV and a porcine internal sample control (ISC) by RT-qPCR was evaluated in serum, oral fluids, and feces. Note: The ISC used in this study is a pig-specific RNA target present in all pig-derived specimens and, therefore, is a measure of sample quality. The RT-qPCR kit used in the study (IDEXX Laboratories, Inc.) detected both PRRSV and ISC RNAs simultaneously.

Serum samples were collected from pigs experimentally inoculated with contemporary wild-type PRRSV-2 (n = 5 pos; n = 5 neg). Oral fluids (6 pos; 4 neg), and feces (5 pos; 5 neg) were from individually housed pigs vaccinated with a PRRSV-2 MLV (Ingelvac® PRRS MLV). Samples were aliquoted into 2 ml tubes to create 4 sets (one set per freeze-thaw "treatment"). Each set was exposed to a specific number of freeze-thaw cycles (2, 5, 10, and 15) by freezing at -80°C and thawing at 4°C overnight. Thereafter, samples (n = 40 per specimen) were tested in technical triplicates for PRRSV and ISC detection using a commercial PRRSV RT-qPCR (IDEXX Laboratories, Inc.). RT-qPCR Cqs were re-expressed as function of the PCR efficiency as "efficiency standardized Cqs (ECqs)" following the equation:  
$$ECq = E^{-?Cq} = E^{-(Cq_{Sample} - Mean Cq \text{ of reference standards})}$$

ECqs were analyzed using a mixed-effects regression model (MRM) in R 4.1.0 (R core team, 2020).

In serum, no freeze-thaw effect was detected in PRRSV and ISC ECqs ( $p > 0.05$ ; MRM). In oral fluids, the slope of the regression line showed a decrease of 0.027 (95% CI: 0.020, 0.033) and 0.052 (95%CI: 0.018, 0.085) ECqs per freeze-thaw cycle for PRRSV and ISC, respectively ( $p < 0.05$ ; MRM). In feces, no freeze-thaw effect was detected for PRRSV or ISC ECqs ( $p > 0.05$ ).

Overall, freeze-thaw cycles had little impact on the detection of PRRSV and the ISC, albeit more so in oral fluids vs serum and feces. Notably, these results apply to PRRSV RNA; further studies are needed to address the effect of freeze-thaw on PCRs for other pathogens and specimen types.